



CERTIFICATE of Conformity

Reference No. : LCS190220001CS

Applicant : Xiamen Weiyou Intelligent Technology Co.,Ltd

Address : The Second East Factory, 215-219 Tian Feng Road ,Xiamen Fujian,
China.

Product : Air Pressure Therapy System

Model(s) : VU-IPC04B

Parameters : For adapter : Input: 100-240VAC, 50/60Hz, 0.8A
Output: 12.6VDC, 2.0A
Air Pressure Therapy System : 12.6VDC, 2.0A

The submitted products have been tested by us with the listed standards and found in compliance with the following European Directives:

the Medical Devices Directive 93/42/EEC

EN 60601-1: 2006+A1: 2013+A12: 2014

The tests were performed in normal operation mode. The test results apply only to the particular sample tested and to the specific tests carried out. This certificate applies specifically to the sample investigated in our test reference number only.

The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.

Other relevant Directives have to be observed.



Date of issue: March 01, 2019



Manager

Guangzhou LCS Compliance Testing Laboratory Ltd.
No 44-1, Qianfeng North Road, Shiqi town, Panyu District, Guangzhou City, China
Tel: (86)020-39166689 Fax: (86)020-39166619
Http://www.LCS-cert.com Email: webmaster@LCS-cert.com



扫码查询真伪
Scan, Query authenticity
1/1