



Attestation of Compliance

Reference No. : LCSA051622144S

Applicant : XIAMEN WEIYOU INTELLIGENT TECHNOLOGY CO.,LTD

Address : UNIT 301 NO. 6 XIANGHONG ROAD,TORCH HI-TECHZONE INDUSTRIAL PARK, XIANG'AN DISTRICT, XIAMEN P.R.CHINA.

Product : Air Pressure Therapy System

Model(s) : VU-IPC12, VU-IPC12C

Parameters : 220-240V~, 50Hz, 95W

Tested according to : EN 60335-1: 2012+A11: 2014+A13: 2017+A1: 2019+A2: 2019+A14: 2019+A15: 2021;
EN 62233: 2008

The submitted products have been tested by us with the listed standards.

This Attestation of Compliance is issued according to The LVD 2014/35/EU, Referred to as the LVD Directive. It confirms that the listed product complies with all essential requirements of the Directive and applies only to the sample and its technical documentation submitted to Shenzhen LCS Compliance Testing Laboratory Ltd. for testing.

After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant Directives have to be observed.



Date of issue: June 13,2022

Manager

